


**CERTIFICATE OF CONFORMITY**  
**(CVG534 Cannery Vacuum Gauge)**

Part Number: CVG534

We hereby certify that the CVG534 Cannery Vacuum Gauge is designed for destructive vacuum sampling testing in canned food and drink products. The non-removable piercing probe is made from type 440C stainless steel and uses a rubber gasket to seal between it and the chrome plated brass housing/fitting. The calibrated accuracy of 0 to 30" Hg copper beryllium bourdon tube is  $\pm 3-2-3\%$  of full scale. The lens is polycarbonate.

**Signed:**



**Samuel P. Schemenauer**  
**(Technician)**

April 24, 2024

**Date**