CHICAGO STAINLESS EQUIPMENT, INC.



1280 S.W. 34th Street • Palm City, Florida • (Phone) 772-781-1441 • (FAX) 772-781-1488 • www.chicagostainless.com

CERTIFICATE OF CONFORMITY

(INSTA-FLOW PRESSURE GAUGE)

Part Number: 2N-E-1/4NPT-GF-BK

Serial Number(s): 23021537

We hereby certify that the product contact surface materials in the above referenced parts are Type 316L stainless steel. The case and ring materials are Type 304 stainless steel

We also hereby certify that the above referenced gauge has been calibrated by direct comparison to master gauges which are periodically referred to one or more of the primary standards traceable to National Institute of Standards & Technology or other national physical measures recognized by NIST. The above referenced gauge has been calibrated and found to be accurate <u>+/- 1½% of full scale</u> at any point within its scale. The accuracy has been certified at every major subdivision on the dial face.

We also hereby certify that the above referenced gauges are tested in the United States of America at our facility in Palm City, Florida.

Signed:

Samuel P. Schemenauer (Technician) February 15, 2023 Date