

Certificate of Calibration

This document certifies that the Tel-Tru Gas Actuate Thermometer described below has been calibrated to its specifications by means of calibration equipment that is current and traceable to the National Institute of Standards and Technology (NIST).

Customer: CHICAGO STAINLESS EQUIP

Tel-Tru Item No.: 41CSCH77BGNAAAA-10

Customer P.O. ID:

C13173

Customer P/N: Sales Order ID:

TU40743

Stem Length:

Temp. Range:

Model No.:

0/1500

CERTIFICATION EQUIPMENT USED:

Description: Certified Glass Thermometer ASTM No.: 65F Serial No.: 3090965

TEMPERATURE CALIBRATION:

Serial No.: 3042079

Setpoint:

65.6C

Checkpoint	Bath Temperature	Actual Reading	Deviation
65-60	65-6C	65.6C	0
10c	100	9.50	0.5
104.40	104.4C	104.5C	0.1

The above Gas Actuate Thermometer has been calibrated to a standard traceable to NIST and the readings are correct to within +/- 1 % full scale, or +/- 1.5

Date Tested: <u>0.2 / 11/13</u>

Calibrator:

QA Engineer:

Tel-Tru Manufacturing Company

MAIN OFFICE: 408 Saint Paul Street, Rochester, New York 14605 MAILING ADDRESS: PO Box 30144, Rochester, New York 14603-3144 USA

Phone: 585.232.1440 • 800.232.5335 • Fax: 585.232.3857 • E-mail: info@teltru.com • Web; www.teltru.com

QUALITY PRODUCTS SINCE 1916

TTF008

Rev. 11/9/12