

CHICAGO STAINLESS EQUIPMENT



1280 S.W. 34th Street • Palm City, Florida • (Phone) 772-781-1441 • (FAX) 772-781-1488 •
www.chicagostainless.com

CERTIFICATE OF CONFORMITY (SANI-FLOW THERMOWELL)

Part Number: TW-2U-4


Serial Number(s): 15040825

We hereby certify that the materials in the above referenced parts are AISA Type 316L stainless steel to the specification in the current revision of ASME BPE, section DT, table DT-3. That the wetted surface has been mechanically polished to a minimum R_a max surface finish of 8 microinches.

The referenced parts are designed and manufactured to conform with 3A SSI sanitary standard for Sensors and Sensor Fittings and Connections 74-xx (Certificate Authorization Number 850).

We also hereby certify that the above referenced instrument(s) are manufactured and/or assembled and tested in the United States of America at our facility in Palm City, Florida.

Signed:



Samuel P. Schemenauer
(Technician)

4/9/2015

Date